

Dual-Color FluoroSpot

Human IFN- γ /IL-17 Kit

Catalog Number ELD5219NL

For the quantitative determination of the frequency of cells releasing human Interferon gamma (IFN- γ) and/or human Interleukin 17 (IL-17).

This package insert must be read in its entirety before using this product.
For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	PAGE
INTRODUCTION	1
PRINCIPLE OF THE ASSAY	1
LIMITATIONS OF THE PROCEDURE	2
TECHNICAL HINTS	2
PRECAUTIONS	2
ASSAY PRINCIPLE	3
MATERIALS PROVIDED & STORAGE CONDITIONS	4
OTHER SUPPLIES REQUIRED	4
SAMPLE PREPARATION	5
REAGENT PREPARATION	5
ASSAY PROCEDURE	6
CALCULATION OF RESULTS	7
REPRODUCIBILITY DATA	7
TROUBLESHOOTING GUIDE	8
REFERENCES	8
PLATE LAYOUT	9

MANUFACTURED AND DISTRIBUTED BY:

USA & Canada | R&D Systems, Inc.

614 McKinley Place NE, Minneapolis, MN 55413, USA
TEL: (800) 343-7475 (612) 379-2956 FAX: (612) 656-4400
E-MAIL: info@RnDSystems.com

DISTRIBUTED BY:

UK & Europe | R&D Systems Europe, Ltd.

19 Barton Lane, Abingdon Science Park, Abingdon OX14 3NB, UK
TEL: +44 (0)1235 529449 FAX: +44 (0)1235 533420
E-MAIL: info@RnDSystems.co.uk

China | R&D Systems China Co., Ltd.

24A1 Hua Min Empire Plaza, 726 West Yan An Road, Shanghai PRC 200050
TEL: +86 (21) 52380373 FAX: +86 (21) 52371001
E-MAIL: info@RnDSystemsChina.com.cn

INTRODUCTION

The Dual-Color Human IFN- γ /IL-17 FluoroSpot assay is designed for the simultaneous detection of human IFN- γ and human IL-17 secreting cells at the single cell level and can be used to simultaneously quantitate the frequency of human IFN- γ and human IL-17 secreting cells. ELISpot assays are well suited for monitoring immune responses to various treatments and therapies and have been used for the quantitation of antigen-specific CD4⁺ and/or CD8⁺ T cell responses. Other methods for assessment of antigen-specific T cell responses, such as chromium release assays with quantitation by limiting dilution, are tedious and require previous *in vitro* expansion of T cells for several days. These assays typically are not suitable for measuring infrequent T cell responses that occur at less than 1 in 1000. ELISpot assays are highly reproducible and sensitive and can be used to measure responses with frequencies well below 1 in 100,000. ELISpot assays do not require prior *in vitro* expansion of T cells and are suitable for high-throughput analysis using only small volumes of primary cells. As such, ELISpot assays are useful tools for research in vaccine development and for the monitoring of various clinical trials.

PRINCIPLE OF THE ASSAY

The enzyme-linked immunospot (ELISpot) assay was originally developed for the detection of individual B cells secreting antigen-specific antibodies (1, 2). This method has since been adapted for the detection of individual cells secreting specific cytokines or other antigens (3, 4). ELISpot assays employ the quantitative sandwich enzyme-linked immunosorbent assay (ELISA) technique.

The Dual-Color FluoroSpot assay provides a method for accurate detection of co-secreted cytokines by the same cell. Instead of enzyme conjugates as in an ELISpot assay, the R&D Systems® novel Dual-Color FluoroSpot assay utilizes NothernLights™ fluorescent probes for the detection of secreted cytokines.

A monoclonal antibody specific for human IFN- γ and a polyclonal antibody specific for human IL-17 are coated onto a polyvinylidene difluoride (PVDF)-backed microplate. The coated microplate is then blocked and appropriately stimulated cells are pipetted into the wells. The microplate is then placed into a humidified 37 °C CO₂ incubator for a specified period of time. During this incubation period, the immobilized antibodies in the immediate vicinity of the secreting cells bind secreted IFN- γ and IL-17. After washing away any cells and unbound substances, a biotinylated polyclonal antibody specific for human IL-17 is added to the wells. Following a wash to remove any unbound antibody, a polyclonal antibody specific for human IFN- γ conjugated to a green fluorophore and streptavidin conjugated to a red fluorophore (for the detection of IL-17) are added to the well. Unbound fluorescent conjugates are subsequently removed by washing and a fluorescence enhancer is added. After incubation with this enhancer, it is removed from the wells, and the microplate is allowed to dry. Green fluorescent spots develop at the sites of IFN- γ secreting cells, and red fluorescent spots appear at the sites of IL-17 secreting cells. The spots can be visualized and counted using either a fluorescent ELISpot reader or using a conventional epifluorescence microscope.

LIMITATIONS OF THE PROCEDURE

- FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
- The kit should not be used beyond the expiration date on the kit label.
- Do not mix or substitute reagents with those from other lots or sources.
- Any variation in pipetting and washing techniques, incubation time or temperature, or kit age can cause variation in the density of spots, intensity of specific staining, and background levels.

TECHNICAL HINTS


- To minimize edge effect, place the microplate (bottom down) onto a piece of aluminum foil (about 4 x 6 inches). Add cells, cover the microplate with the lid, and shape the foil around the edges of the microplate. The foil may be left on the microplate until after the Fluorescence Enhancer has been aspirated.
- Do not remove the flexible plastic underdrain on the bottom of the microplate before or during incubation and development. It may damage the PVDF membrane filter. The underdrain cover may be removed only after completing the incubation with the Fluorescence Enhancer.
- To avoid damage to the membrane, do not touch PVDF membrane filters with pipette tips when pipetting cells and reagents.
- Upon completing the experiment, do not dry the microplate at a temperature above 37 °C. It may cause the PVDF membrane filters to crack.
- The 96-well microplate provided in this kit is not sterile. Due to the short incubation period and the presence of antibiotics in the culture media, microbial contamination has not been shown to be an issue with this FluoroSpot procedure.
- This kit is designed for single use only. The layout of the assay should be carefully planned to maximize the use of the provided microplate and reagents.
- The controls listed are recommended for each FluoroSpot experiment:
 - Positive Control** - Use recombinant human IFN- γ and recombinant human IL-17 in separate wells.
 - Unstimulated/Negative Control** - Use the same number of unstimulated cells as stimulated cells.
 - Background Control** - Use sterile culture media.
 - Detection Antibody Control** - Substitute phosphate buffered saline for detection antibody.

PRECAUTIONS


Some components of this kit contain sodium azide, which may react with lead and copper plumbing to form explosive metallic azides. Flush with large volumes of water during disposal. Some components in this kit contain a preservative which may cause an allergic skin reaction. Avoid breathing mist.

Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling. Refer to the MSDS on our website prior to use.

ASSAY PRINCIPLE



Two different proteins secreted by cultured cells bind to the capture antibodies coated on the well.




Remove cells by washing.
Add the biotinylated detection antibody which binds to the FIRST captured protein.


Remove the unbound biotinylated detection antibody by washing.

Add the Streptavidin conjugated to NL557 fluorescent tag (red).

Add the detection antibody conjugated to A488 fluorescent tag (green) which binds to the SECOND captured protein.










After incubation with the NorthernLights™ fluorescent tags, remove the unbound fluorescent tags by washing and add the fluorescent enhancer detection mixture. After incubation with this enhancers, aspirate the solution and allow microplate to dry. Analyze using either an ELISpot reader capable of detecting green and red fluorescent spots or conventional epifluorescence microscope.



Red fluorescent spots represent the FIRST captured protein and green fluorescent spots represent the SECOND captured protein.

KEY

-  FIRST secreted protein
-  SECOND secreted protein
-  Capture antibody against the FIRST secreted protein
-  Capture antibody against the SECOND secreted protein
-  Biotin-conjugated detection antibody against the FIRST protein
-  A488-conjugated detection antibody against the SECOND protein
-  Streptavidin conjugated to NL557 fluorescent tag

MATERIALS PROVIDED & STORAGE CONDITIONS

Store the unopened kit at 2-8 °C. Do not use past kit expiration date.

Note: *This kit is validated for single use only. Results obtained using previously opened or reconstituted reagents may not be reliable.*

PART	PART #	DESCRIPTION
Microplate	607760	96-well PVDF-backed microplate used for fluorescent applications.
Human IL-17 Capture Ab Concentrate	894261	150 µL of a concentrated solution of polyclonal antibody specific for human IL-17.
Human IFN-γ Capture Ab Concentrate	894237	150 µL of a concentrated solution of monoclonal antibody specific for human IFN-γ.
Human IL-17 Detection Ab Concentrate	894803	150 µL of a concentrated solution of biotinylated polyclonal antibody specific for human IL-17 with preservatives.
Human IFN-γ Detection Ab Concentrate A488	894238	250 µL of a concentrated solution of fluorescent green conjugated polyclonal antibody specific for human IFN-γ with preservatives.
Streptavidin-NL557 Concentrate	967367	50 µL of Streptavidin conjugated to NorthernLights™ NL557 with preservatives.
Dilution Buffer 1	895307	12 mL of a buffer for diluting Human IL-17 Detection Ab Concentrate with preservatives.
Dilution Buffer 2	895354	12 mL of a buffer for diluting Human IFN-γ Detection Ab A488 and Streptavidin-NL557 Concentrates with preservatives.
Wash Buffer 10X Concentrate	895308	50 mL of a concentrated solution of a buffered surfactant with preservative.
Fluorescence Enhancer	894236	12 mL of a solution to enhance green NorthernLights™ fluorescence.
Human IL-17 Positive Control	892729	3 ng of recombinant human IL-17 with preservatives; lyophilized.
Human IFN-γ Positive Control	890893	4 ng of recombinant human IFN-γ with preservatives; lyophilized.

OTHER SUPPLIES REQUIRED

- ELISpot reader capable of detecting fluorescence or a conventional epifluorescence microscope.
- Pipettes and pipette tips.
- 35% Ethanol or Methanol (refer to MSDS prior to use).
- Deionized water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- PBS (137 mM NaCl, 2.7 mM KCl, 8.1 mM Na₂HPO₄, 1.5 mM KH₂PO₄, pH 7.2-7.4, 0.2 µm filtered, R&D Systems®, Catalog #DY006).
- Block Buffer (1% BSA and 5% sucrose in PBS).
- 500 mL graduated cylinder.
- 37 °C CO₂ incubator.
- Sterile culture media.

SAMPLE PREPARATION

The types of effector and responder cells used, method of cell separation, mode of stimulation, and length of incubation are to be determined by each investigator.

REAGENT PREPARATION

Bring all reagents to room temperature, except the detection antibody concentrates and Dilution Buffer 1, which should remain at 2-8 °C.

1X Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. To prepare Wash Buffer, add 50 mL of Wash Buffer Concentrate to 450 mL of deionized water and mix well.

Positive Controls - Reconstitute the lyophilized Human IL-17 Positive Control and Human IFN- γ Positive Control with 250 μ L of culture medium that is used to incubate cells.

Capture Antibody Mixture (Human IL-17 + Human IFN- γ) - Tap or vortex each vial to release reagent collected in the cap. Transfer 100 μ L of Human IL-17 Capture Ab Concentrate and 100 μ L of Human IFN- γ Capture Ab Concentrate into 12 mL of PBS and mix well. **For optimal performance, prepare the Capture Antibody Mixture immediately before use.**

Human IL-17 Detection Antibody - Tap or vortex the vial to release reagent collected in the cap. Transfer 100 μ L of Human IL-17 Detection Ab Concentrate into the vial labeled Dilution Buffer 1 and mix well. **For optimal performance, prepare the Human IL-17 Detection Antibody immediately before use.**

Human IFN- γ Detection Antibody A488 + Streptavidin-NL557 Mixture - Tap or vortex each vial to release reagent collected in the cap. Transfer 200 μ L of Human IFN- γ Detection Ab Concentrate A488 and 24 μ L of Streptavidin-NL557 Concentrate into the vial labeled Dilution Buffer 2 and mix well. **For optimal performance, prepare the Human IFN- γ Detection Antibody A488 + Streptavidin-NL557 Mixture immediately before use.**

ASSAY PROCEDURE

Bring all reagents to room temperature, except the Capture Antibody Mixture, Human IL-17 Detection Antibody, and Human IFN- γ Detection Antibody A488 + Streptavidin-NL557 Mixture, which should remain at 2-8 °C. All samples and controls should be assayed at least in duplicate. A plate layout is provided to record controls and samples assayed.

1. Prepare the membranes by adding 15 μ L of 35% alcohol to each well. Incubate for 1 minute.
2. Aspirate each well and wash, repeating the process four times for a total of five washes. Wash by filling each well with deionized water (250-300 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining liquid by aspirating or decanting. Invert the plate and blot it against clean paper towels. **Note:** *Adjust the height of the manifold dispenser or autowasher to prevent damage to the membranes.*
3. Add 100 μ L of the diluted Capture Antibody Mixture into each well, and incubate overnight at 2-8 °C.
4. Aspirate each well and wash **with 1X PBS** (as in step 2), repeating the process three times for a total of four washes.
5. Fill each well of the microplate with 200 μ L of Block Buffer, and incubate the microplate for approximately 90 minutes at room temperature.
6. Aspirate the Block Buffer, and fill each well of the microplate with 200 μ L of sterile culture media. Incubate for approximately 20 minutes at room temperature.
7. When the cells are ready to be plated, aspirate the culture media from the wells. Immediately add 100 μ L of the appropriate cells or controls to each well. Refer to the Technical Hints section for appropriate controls.
8. Incubate the cells in a humidified 37 °C CO₂ incubator. Optimal incubation time for each stimulus should be determined by the investigator. **Do not disturb the cells during the incubation period.**
9. Aspirate each well and wash **with Wash Buffer** (as in step 2), repeating the process three times for a total of four washes.
10. Add 100 μ L of the diluted Human IL-17 Detection Antibody to each well, and incubate overnight at 2-8 °C.
11. Repeat the wash procedure described in step 9.
12. Add 100 μ L of the diluted Human IFN- γ Detection Antibody A488 + Streptavidin-NL557 Mixture to each well, and incubate for 2 hours at room temperature. **Protect from light.**
13. Repeat the wash procedure described in step 9 **using 1X PBS instead of Wash Buffer.**
14. Add 100 μ L of the Fluorescence Enhancer Solution to each well, and incubate for 15 minutes at room temperature. **Protect from light.**
15. Aspirate the Fluorescence Enhancer Solution. Remove the underdrain from the back of the microplate, and rinse with deionized water. **Do not rinse the wells.** Let the microplate dry completely before analyzing. **Protect from light until analysis.**

CALCULATION OF RESULTS

The developed microplate can be analyzed by counting spots using either an ELISpot reader capable of detecting green and red fluorescent spots or a conventional epifluorescence microscope. Specific spots are round and have a dark center with slightly fuzzy edges. Quantification of results can be done, for example, by calculating the number of spot forming cells (SFC) per number of cells added to the well.

REPRODUCIBILITY DATA

Peripheral blood mononuclear cells (2×10^6 cells/mL) were stimulated with 50 ng/mL of phorbol 12-myristate-13-acetate, 0.5 $\mu\text{g/mL}$ of calcium ionomycin, 0.1 $\mu\text{g/mL}$ of anti-CD28, and 0.5 $\mu\text{g/mL}$ of anti-CD3 ϵ overnight at 37 °C in a 5% CO₂ incubator. The sample was assayed in five wells according to the procedure and analyzed with a dissection microscope.

Well	Number of Green (IFN- γ) Spots Counted	Number of Red (IL-17) Spots Counted
1	332	28
2	373	32
3	329	36
4	364	29
5	359	33

TROUBLESHOOTING GUIDE

Observation	Problem	Corrective Action
Following the incubation with the Fluorescence Enhancer and rinsing the back of the microplate with deionized water, the green (or red) background color of the filter membrane attenuates visualization and quantification of spots.	The membrane is wet.	Microplates cannot be analyzed accurately until the PVDF filter membranes are completely dry. Wait until the membrane becomes dry (typically 15-30 minutes at 37 °C or 60-90 minutes at room temperature).
The number of spots in the wells that contained the cells is high, but their contrast as well as intensity of staining in the Positive Control wells is low.	Underdevelopment; perhaps the result of using Streptavidin-NL557 and/or Fluorescence Enhancer solutions that have not been brought to room temperature.	Warm the appropriate reagents to room temperature before adding them to the wells.
The number of spots in the wells that contained cells is lower than expected whereas Positive Control wells turned green (or red).	Cell stimulation problem.	Ensure that reagents used to stimulate the cytokine release from the cells retained their biological activity. One way to check is to perform immunohistochemistry on fixed cells after stimulation.
	Too few cells were added to the wells.	Increase the number of cells added per well.
Following incubation with the Fluorescence Enhancer and drying the microplate, the density of the spots makes it difficult to quantify.	Too many cells were added to the wells.	Make dilutions of cells (1×10^6 , 5×10^5 , 1×10^5 , 5×10^4 , 1×10^4 cells per well) to determine the optimal number of cells that will result in formation of distinct spots.

REFERENCES

1. Czerkinsky, C.C. *et al.* (1983) J. Immunol. Methods **65**:109.
2. Sedgwick, J.D. and P.G. Holt (1983) J. Immunol. Methods **57**:301.
3. Czerkinsky, C.C. *et al.* (1984) J. Immunol. Methods **72**:489.
4. Helms, T. *et al.* (2000) J. Immunol. **164**:3723.

PLATE LAYOUT

Use this plate layout to record controls and samples assayed.

12								
11								
10								
9								
8								
7								
6								
5								
4								
3								
2								
1								
	A	B	C	D	E	F	G	H

All trademarks and registered trademarks are the property of their respective owners.