

Quantikine[®] ELISA

Mouse Erythropoietin Immunoassay

Catalog Number MEP00B

For the quantitative determination of mouse Erythropoietin (Epo) concentrations in cell culture supernates, tissue homogenates, serum, and plasma.

This package insert must be read in its entirety before using this product.
For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION	PAGE
INTRODUCTION	1
PRINCIPLE OF THE ASSAY	2
LIMITATIONS OF THE PROCEDURE	2
TECHNICAL HINTS	2
MATERIALS PROVIDED & STORAGE CONDITIONS	3
OTHER SUPPLIES REQUIRED	3
PRECAUTIONS	4
SAMPLE COLLECTION & STORAGE	4
SAMPLE PREPARATION	4
REAGENT PREPARATION	5
ASSAY PROCEDURE	6
CALCULATION OF RESULTS	7
TYPICAL DATA	7
PRECISION	8
RECOVERY	8
LINEARITY	8
SENSITIVITY	9
CALIBRATION	9
SAMPLE VALUES	9
SPECIFICITY	10
REFERENCES	11
PLATE LAYOUT	12

MANUFACTURED AND DISTRIBUTED BY:

USA & Canada | R&D Systems, Inc.

614 McKinley Place NE, Minneapolis, MN 55413, USA
TEL: (800) 343-7475 (612) 379-2956 FAX: (612) 656-4400
E-MAIL: info@RnDSystems.com

DISTRIBUTED BY:

UK & Europe | R&D Systems Europe, Ltd.

19 Barton Lane, Abingdon Science Park, Abingdon OX14 3NB, UK
TEL: +44 (0)1235 529449 FAX: +44 (0)1235 533420
E-MAIL: info@RnDSystems.co.uk

China | R&D Systems China Co., Ltd.

24A1 Hua Min Empire Plaza, 726 West Yan An Road, Shanghai PRC 200050
TEL: +86 (21) 52380373 FAX: +86 (21) 52371001
E-MAIL: info@RnDSystemsChina.com.cn

INTRODUCTION

Erythropoietin (Epo) is a 34-39 kDa secreted glycoprotein that is a member of the type I cytokine superfamily. The mouse Epo gene encodes a 192 amino acid (aa) residue precursor that contains a 26 aa signal peptide and a 166 aa mature protein containing three potential N-linked glycosylation sites (1-4). Mouse Epo lacks the O-linked glycosylation site found in human Epo. Although carbohydrate chains are not required for *in vitro* receptor binding, they are required for *in vivo* Epo bioactivity. Depending on the cell source, different Epo isoforms are produced that differ in their glycan compositions and sialic acid contents (5-8). Mature mouse and rat Epo share 94% aa sequence identity. They also share from 80%-82% aa identity with mature human, porcine, rhesus monkey and feline Epo (2, 3). Epo is primarily produced by cells in the kidney (interstitial peritubular renal fibroblasts) and liver (hepatocytes and Ito cells), where its production is up-regulated by hypoxia. Other tissues and cells, including neural tissues (astrocytes and neurons), testis (Sertoli cells), uterus, placenta, and erythroid progenitors, have also been shown to produce Epo (9-14).

Epo is best known for its role in red blood cell formation. While Epo is not a lineage commitment factor, it inhibits apoptosis and induces burst forming unit-erythroid (BFU-E) differentiation into colony forming unit-erythroid (CFU-E), and the subsequent proliferation and maturation of CFU-E into early normoblasts (10, 15, 16). Apart from its role in erythropoiesis, Epo also acts on various non-hematopoietic cells to function as a viability and proliferation factor. Epo can stimulate myoblast proliferation while suppressing its differentiation, resulting in the expansion of the progenitor cell population (17). Epo is a tissue-protective factor that protects against ischemic and toxic injuries to neuronal, cardiovascular and renal tissues (18, 19). Epo has also been shown to promote angiogenesis in various physiologic and pathologic conditions (20, 21).

Epo binds and signals via the high-affinity preformed homodimeric Epo receptor (Epo R) that is composed of two Epo R subunits. Each Epo R subunit is a type I transmembrane glycoprotein that belongs to the type I cytokine receptor superfamily (18, 22-24). Its extracellular domain contains the characteristic two fibronectin type III domains and a WSxWS motif near the plasma membrane (24, 25). Binding of Epo to the Epo R homodimer results in conformational change and phosphorylation and activation of the non-receptor protein kinase JAK2, which activates the downstream signaling cascade (26). An alternative Epo heteromeric receptor complex that transduces cell-protective signals and containing the β common receptor (β CR) subunit in addition to the Epo R subunit has been described. β CR also belongs to the type I cytokine receptor superfamily and is a subunit that is shared by the heteromeric IL-3, IL-5 and GM-CSF receptor complexes. Epo binds with lower affinity to the heteromeric receptor consisting of a Epo R subunit and a β CR homodimer (18).

The Quantikine® Mouse Erythropoietin Immunoassay is a 4.5 hour solid-phase ELISA designed to measure mouse Epo in cell culture supernates, tissue homogenates, serum, and plasma. It contains antibodies raised against recombinant Epo and has been shown to accurately quantitate the recombinant factor. Results obtained using natural mouse Epo showed linear curves that were parallel to the standard curves obtained using the Quantikine® kit standards. These results indicate that this kit can be used to determine relative mass values for natural mouse Epo.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for mouse Epo has been pre-coated onto a microplate. Standards, control, and samples are pipetted into the wells and any Epo present is bound by the immobilized antibody. After washing away any unbound substances, an enzyme-linked monoclonal antibody specific for mouse Epo is added to the wells. Following a wash to remove any unbound antibody-enzyme reagent, a substrate solution is added to the wells. The enzyme reaction yields a blue product that turns yellow when the Stop Solution is added. The intensity of the color measured is in proportion to the amount of Epo bound in the initial step. The sample values are then read off the standard curve.

LIMITATIONS OF THE PROCEDURE

- FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
- The kit should not be used beyond the expiration date on the kit label.
- Do not mix or substitute reagents with those from other lots or sources.
- If samples generate values higher than the highest standard, further dilute the samples with calibrator diluent and repeat the assay.
- Any variation in diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.
- Variations in sample collection, processing, and storage may cause sample value differences.
- This assay is designed to eliminate interference by other factors present in biological samples. Until all factors have been tested in the Quantikine® Immunoassay, the possibility of interference cannot be excluded.

TECHNICAL HINTS

- When mixing or reconstituting protein solutions, always avoid foaming.
- To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Substrate Solution should remain colorless until added to the plate. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
- Stop Solution should be added to the plate in the same order as the Substrate Solution. The color developed in the wells will turn from blue to yellow upon addition of the Stop Solution.

MATERIALS PROVIDED & STORAGE CONDITIONS

Store the unopened kit at 2-8 °C. Do not use past kit expiration date.

PART	PART #	DESCRIPTION	STORAGE OF OPENED/ RECONSTITUTED MATERIAL
Mouse Epo Microplate	892567	96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody specific for mouse Epo.	Return unused wells to the foil pouch containing the desiccant pack. Reseal along entire edge of zip-seal. May be stored for up to 1 month at 2-8 °C.*
Mouse Epo Standard	892569	Recombinant Epo in a buffered protein base with preservatives; lyophilized. <i>Refer to the vial label for reconstitution volume.</i>	Aliquot and store for up to 1 month at ≤ -20 °C in a manual defrost freezer.* Avoid repeated freeze-thaw cycles.
Mouse Epo Control	892570	Recombinant Epo in a buffered protein base with preservatives; lyophilized. The assay value of the control should be within the range specified on the label.	
Mouse Epo Conjugate	892568	12 mL of a monoclonal antibody specific for mouse Epo conjugated to horseradish peroxidase with preservatives.	May be stored for up to 1 month at 2-8 °C.*
Assay Diluent RD1W	895038	12 mL of a buffered protein solution with preservatives.	
Calibrator Diluent RD6Z	895466	21 mL of a buffered protein solution with preservatives.	
Wash Buffer Concentrate	895003	21 mL of a 25-fold concentrated solution of buffered surfactant with preservative. <i>May turn yellow over time.</i>	
Color Reagent A	895000	12 mL of stabilized hydrogen peroxide.	
Color Reagent B	895001	12 mL of stabilized chromogen (tetramethylbenzidine).	
Stop Solution	895174	23 mL of diluted hydrochloric acid.	
Plate Sealers	N/A	4 adhesive strips.	

* Provided this is within the expiration date of the kit.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 500 mL graduated cylinder.
- Horizontal orbital microplate shaker (0.12" orbit) capable of maintaining a speed of 500 ± 50 rpm.
- Test tubes for dilution of standards and samples.

PRECAUTIONS

The Stop Solution provided with this kit is an acid solution.

Some components in this kit contain preservative which may cause an allergic skin reaction. Avoid breathing mist.

Color Reagent B may cause skin, eye, and respiratory irritation. Avoid breathing fumes.

Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling. Refer to the MSDS on our website prior to use.

SAMPLE COLLECTION & STORAGE

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤ -20 °C. Avoid repeated freeze-thaw cycles.

Tissue Homogenates - The preparation of tissue homogenates will vary depending upon the tissue type. For this assay, heart, lung, and spleen tissue from three mice was rinsed with 1X PBS to remove excess blood, homogenized in 5-10 mL of 1X PBS, and stored overnight at ≤ -20 °C. After two freeze-thaw cycles to break up the cell membranes, the homogenates were centrifuged for 5 minutes at 5000 x g. Homogenates should be assayed immediately or aliquotted and stored at ≤ -20 °C. Avoid repeated freeze-thaw cycles.

Serum - Allow blood samples to clot for 2 hours at room temperature before centrifuging for 20 minutes at 2000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20 °C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using heparin as an anticoagulant. Centrifuge for 20 minutes at 2000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20 °C. Avoid repeated freeze-thaw cycles.

Note: *EDTA and citrate plasma have not been validated for use in this assay.*

Grossly hemolyzed or lipemic samples may not be suitable for use in this assay.

SAMPLE PREPARATION

Serum and plasma samples require a 2-fold dilution into Calibrator Diluent RD6Z prior to assay. A suggested 2-fold dilution is 70 μ L of sample + 70 μ L of Calibrator Diluent RD6Z.

REAGENT PREPARATION

Bring all reagents to room temperature before use.


Mouse Epo Control - Reconstitute the control with 1.0 mL of deionized or distilled water. Mix thoroughly. Assay the control undiluted.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 20 mL Wash Buffer Concentrate to deionized or distilled water to prepare 500 mL of Wash Buffer.

Substrate Solution - Color Reagents A and B should be mixed together in equal volumes within 15 minutes of use. Protect from light. 100 μ L of the resultant mixture is required per well.

Mouse Epo Standard - Refer to the vial label for reconstitution volume. Reconstitute the Mouse Epo Standard with Calibrator Diluent RD6Z. Do not substitute other diluents. This reconstitution produces a stock solution of 3000 pg/mL. Allow the stock solution to sit for a minimum of 5 minutes with gentle mixing prior to making dilutions.

Pipette 200 μ L of Calibrator Diluent RD6Z into each tube. Use the stock solution to produce a dilution series (below). Mix each tube gently but thoroughly before the next transfer. The undiluted Mouse Epo Standard (3000 pg/mL) serves as the high standard. Calibrator Diluent RD6Z serves as the zero standard (0 pg/mL).



ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that all samples, control, and standards be assayed in duplicate.

1. Prepare all reagents, standard dilutions, control, and samples as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the foil pouch containing the desiccant pack, and reseal.
3. Add 50 μ L of Assay Diluent RD1W to each well.
4. Add 50 μ L of standard, control, or sample* per well. Cover with the adhesive strip provided. Incubate for 2 hours at room temperature on a horizontal orbital microplate shaker (0.12" orbit) set at 500 ± 50 rpm. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process four times for a total of five washes. Wash by filling each well with Wash Buffer (400 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 μ L of Mouse Epo Conjugate to each well. Cover with a new adhesive strip. Incubate for 2 hours at room temperature on the shaker.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 μ L of Substrate Solution to each well. Incubate for 30 minutes at room temperature **on the benchtop. Protect from light.**
9. Add 100 μ L of Stop Solution to each well. Gently tap the plate to ensure thorough mixing.
10. Determine the optical density of each well within 30 minutes, using a microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

*Samples may require dilution. See Sample Preparation section.

CALCULATION OF RESULTS


Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density (O.D.).

Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the mouse Epo concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

TYPICAL DATA

This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.



(pg/mL)	O.D.	Average	Corrected
0	0.041 0.045	0.043	—
47	0.078 0.082	0.080	0.037
94	0.113 0.118	0.116	0.073
188	0.187 0.192	0.190	0.147
375	0.339 0.345	0.342	0.299
750	0.658 0.660	0.659	0.616
1500	1.284 1.314	1.299	1.256
3000	2.473 2.521	2.497	2.454

PRECISION

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Three samples of known concentration were tested in twenty separate assays to assess inter-assay precision. Assays were performed by at least three technicians using two lots of components.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean (pg/mL)	196	258	704	180	239	700
Standard deviation	7.7	7.8	31.2	17.5	15.3	17.5
CV (%)	3.9	3.0	4.4	9.7	6.4	2.5

RECOVERY

The recovery of mouse Epo spiked to three levels throughout the range of the assay in various matrices was evaluated.

Sample Type	Average % Recovery	Range
Cell culture supernates (n=7)	110	102-117%
Tissue homogenates (n=3)	94	80-111%
Serum* (n=9)	93	85-114%
Heparin plasma* (n=8)	103	90-120%

*Serum and plasma samples were spiked and then diluted as described in the Sample Preparation section.

LINEARITY

To assess the linearity of the assay, samples containing or spiked with high concentrations of mouse Epo in each matrix were diluted with calibrator diluent and then assayed.

		Cell culture supernates (n=9)	Tissue homogenates (n=3)	Serum* (n=6)	Heparin plasma* (n=9)
1:2	Average % of Expected	92	106	95	94
	Range (%)	90-96	94-120	91-98	89-97
1:4	Average % of Expected	93	105	96	96
	Range (%)	90-102	92-119	90-100	88-102
1:8	Average % of Expected	94	107	98	97
	Range (%)	90-100	94-120	90-105	87-106
1:16	Average % of Expected	95	106	106	97
	Range (%)	85-103	95-113	95-117	83-108

*Serum and plasma samples were spiked and then diluted as described in the Sample Preparation section.

SENSITIVITY

Nineteen assays were evaluated and the minimum detectable dose (MDD) of mouse Epo ranged from 6.5-46.9 pg/mL. The mean MDD was 18.0 pg/mL.

The MDD was determined by adding two standard deviations to the mean O.D. value of twenty zero standard replicates and calculating the corresponding concentration.

CALIBRATION

This immunoassay is calibrated against a highly purified *Sf 21*-expressed recombinant mouse Epo produced at R&D Systems®.

SAMPLE VALUES

Serum/Plasma - Twenty mouse samples were evaluated for the presence of mouse Epo in this assay.

NSA Mouse Strain	Mean of Detectable (pg/mL)	% Detectable	Range (pg/mL)
Serum (n=20)	237	65	ND-683
Heparin plasma (n=20)	151	30	ND-198

ND=Non-detectable

Tissue Homogenates - Lung, heart, and spleen tissues from three adult female mice were prepared as described in the Sample Collection and Storage section on page 4. Supernates were removed, tested for mouse Epo, and measured 61 pg/mL, 111 pg/mL, and 75 pg/mL, respectively.

SPECIFICITY

This assay recognizes natural and recombinant mouse Epo.

The factors listed below were prepared at 50 ng/mL in calibrator diluent and assayed for cross-reactivity. Preparations of the following factors at 50 ng/mL in a mid-range recombinant mouse Epo control were assayed for interference. No significant cross-reactivity or interference was observed.

Recombinant mouse:

ICAM-1	IL-6	Leptin
IFN- γ	IL-7	LIF
IGF-I	IL-9	M-CSF
IGF-II	IL-10	MIP-1 α
IL-1 α	IL-11	MIP-1 β
IL-1 β	IL-12	OSM
IL-1ra	IL-12 p40	RANTES
IL-2	IL-13	TNF RII
IL-3	IL-17	TNF- α
IL-4	IL-18	Tpo
IL-5	JE/MCP-1	VEGF

Recombinant mouse Epo R/Fc Chimera interferes at concentrations > 2.0 ng/mL in this assay.

Recombinant human Epo has approximately 9% cross-reactivity in this assay at concentrations \geq 2.0 ng/mL.


Recombinant rat Epo has 100% cross-reactivity in this assay. This assay is not validated for use with rat samples because rat sample dilution is not linear.

REFERENCES

1. Nagao, M. *et al.* (1992) *Biochem. Biophys. Acta* **1171**:99.
2. Wen, D. *et al.* (1993) *Blood* **82**:1507.
3. Jacobs, K. *et al.* (1985) *Nature* **313**:806.
4. Shoemaker, C.B. and L.D. Mitssock (1986) *Mol. Cell. Biol.* **6**:849.
5. Takeuchi, M. *et al.* (1988) *J. Biol. Chem.* **263**:3657.
6. Takeuchi, M. and A. Kobata (1991) *Glycobiology* **1**:337.
7. Lai, P-H. *et al.* (1986) *J. Biol. Chem.* **261**:3116.
8. Storrington, P.L. *et al.* (1998) *Br. J. Haematol.* **100**:79.
9. Fisher, J.W. *et al.* (1996) *Br. J. Haematol.* **95**:27.
10. Lacombe, C. (1997) *Eur. Cytokine Netw.* **8**:308.
11. Gopfert, T. *et al.* (1996) *J. Cell. Physiol.* **168**:354.
12. Magnanti, M. *et al.* (2001) *Blood* **98**:2872.
13. Stopka, T. *et al.* (1998) *Blood* **91**:3766.
14. Bernaudin, M. *et al.* (2000) *Glia* **30**:271.
15. Muta, K. *et al.* (1994) *J. Clin. Invest.* **94**:34.
16. Fisher, J.W. (2003) *Exp. Biol. Med.* **228**:1.
17. Ogilvie, M. *et al.* (2000) *J. Biol. Chem.* **275**:39754.
18. Brines, M. and A. Cerami (2005) *Nat. Rev. Neurosci.* **6**:484.
19. Chong, Z.Z. *et al.* (2002) *J. Cereb. Blood Flow Metab.* **22**:503.
20. Ribatti, D. *et al.* (1999) *Blood* **93**:2627.
21. Haroon, Z.A. *et al.* (2003) *Am. J. Pathol.* **163**:993.
22. Masuda, S. *et al.* (1993) *J. Biol. Chem.* **268**:11208.
23. D'Andrea, A.D. *et al.* (1989) *Cell* **57**:277.
24. Wojchowski, D.M. *et al.* (1999) *Exp. Cell Res.* **253**:143.
25. Furmanek, A. *et al.* (2003) *Biochemistry* **42**:8452.
26. Watowich, S.S. *et al.* (1999) *Int. J. Biochem. Cell Biol.* **31**:1075.

PLATE LAYOUT

Use this plate layout to record standards and samples assayed.



NOTES

NOTES

All trademarks and registered trademarks are the property of their respective owners.

©2017 R&D Systems®, Inc.