Quantikine[®] ELISA

Mouse Fas Ligand/TNFSF6 Immunoassay

Catalog Number MFL00

For the quantitative determination of mouse Fas Ligand concentrations in cell culture supernates, serum, and plasma.

This package insert must be read in its entirety before using this product. For research use only. Not for use in diagnostic procedures.

TABLE OF CONTENTS

SECTION

PAGE

INTRODUCTION1	
PRINCIPLE OF THE ASSAY2)
LIMITATIONS OF THE PROCEDURE)
TECHNICAL HINTS)
MATERIALS PROVIDED & STORAGE CONDITIONS	;
OTHER SUPPLIES REQUIRED	ŀ
PRECAUTIONS4	ŀ
SAMPLE COLLECTION & STORAGE	ŀ
REAGENT PREPARATION	;
ASSAY PROCEDURE	;
CALCULATION OF RESULTS	,
TYPICAL DATA	,
PRECISION8	3
RECOVERY	3
LINEARITY	;
SENSITIVITY)
CALIBRATION)
SAMPLE VALUES)
SPECIFICITY)
REFERENCES	
PLATE LAYOUT)

MANUFACTURED AND DISTRIBUTED BY:

USA & Canada | R&D Systems, Inc.

614 McKinley Place NE, Minneapolis, MN 55413, USA TEL: (800) 343-7475 (612) 379-2956 FAX: (612) 656-4400 E-MAIL: info@RnDSystems.com

DISTRIBUTED BY:

UK & Europe | R&D Systems Europe, Ltd.

19 Barton Lane, Abingdon Science Park, Abingdon OX14 3NB, UK TEL: +44 (0)1235 529449 FAX: +44 (0)1235 533420 E-MAIL: info@RnDSystems.co.uk

China | R&D Systems China Co., Ltd.

24A1 Hua Min Empire Plaza, 726 West Yan An Road, Shanghai PRC 200050 TEL: +86 (21) 52380373 FAX: +86 (21) 52371001 E-MAIL: info@RnDSystemsChina.com.cn

INTRODUCTION

Mouse Fas Ligand is a member of the TNF superfamily of molecules that induces apoptosis in susceptible cells expressing the fibroblast associated (Fas) receptor (1). Mouse Fas Ligand is expressed on a number of cell types, including endothelial cells (2), macrophages (3), CD4⁺ and CD8⁺ T cells (4, 5), neutrophils (6, 7), Sertoli cells and Leydig cells (8), thyroid epithelium (9, 10), fibroblasts (11), microglia (12), corneal epithelium (13), renal proximal epithelium (14), monocytes (15), mammary epithelium (16), and myeloid dendritic cells (17, 18). The cDNA for mouse Fas Ligand codes for a 279 amino acid (aa) residue, 40 kDa type II transmembrane protein that contains a 179 aa extracellular region, a 22 aa transmembrane segment, and a 78 aa cytoplasmic domain (19-23). A point mutation at residue 273 of the extracellular domain results in a molecule with a generalized lymphoproliferative disease (gld) phenotype (20, 23). Membrane-bound Fas Ligand can be cleaved by metalloproteinases to produce a 150 aa, 26 kDa soluble form that circulates as a nondisulfide-linked trimer (19-25). An alternate splice form also exists that is 16 kDa, 69 aa long, and constitutively expressed. This form is nonapoptotic, as it blocks anti-Fas activation (26). Both membrane-bound and soluble Fas Ligand are bioactive, with the highest activity associated with membrane-bound Fas Ligand (24, 27). Both mouse and human Fas Ligand are active on mouse and human cells (28). Mouse soluble Fas Ligand shares 94% and 85% aa identity to rat and human Fas Ligand, respectively (28, 29).

The signaling receptor for Fas Ligand is Fas, a 45 kDa type I transmembrane glycoprotein that is 245 aa in length (30). There is also a second reported receptor for Fas Ligand termed DcR3 (decoy receptor 3) which is known only in soluble form (31). Both Fas and DcR3 bind Fas Ligand with equal affinity (31).

Fas-mediated apoptosis plays a critical role during development, in liver hemostasis, and in immunological reactions. In the immune system, Fas Ligand is suggested to play an important role in T cell activation. During an immune response, naive T cells are exposed to antigens. This leads to T cell proliferation and differentiation. To insure that proliferation is self-limiting, Fas Ligand is upregulated on activated T cells and interacts with Fas on either the same cell or bystander cells such as unstimulated T cells and macrophages. Fas-induced apoptosis of these cells interrupts the cycle of cytokine stimulation and proliferation (19, 32, 33). In addition to activating Fas, Fas Ligand can also be activated by Fas, transmitting a reverse signal that results in proliferation of CD4⁺ naive T cells (5, 33).

The Quantikine[®] Mouse Fas Ligand/TNFSF6 Immunoassay is a 4.5 hour solid-phase ELISA designed to measure mouse Fas Ligand in cell culture supernates, serum, and plasma. It contains NS0-expressed recombinant mouse Fas Ligand and antibodies raised against the recombinant factor. This immunoassay has been shown to accurately quantitate the recombinant factor. Results obtained using natural mouse Fas Ligand showed linear curves that were parallel to the standard curves obtained using the Quantikine[®] kit standards. These results indicate that this kit can be used to determine relative mass values for natural mouse Fas Ligand.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for mouse Fas Ligand has been pre-coated onto a microplate. Standards, control, and samples are pipetted into the wells and any Fas Ligand present is bound by the immobilized antibody. After washing away any unbound substances, an enzyme-linked polyclonal antibody specific for mouse Fas Ligand is added to the wells. Following a wash to remove any unbound antibody-enzyme reagent, a substrate solution is added to the wells. The enzyme reaction yields a blue product that turns yellow when the Stop Solution is added. The intensity of the color measured is in proportion to the amount of Fas Ligand bound in the initial step. The sample values are then read off the standard curve.

LIMITATIONS OF THE PROCEDURE

- FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
- The kit should not be used beyond the expiration date on the kit label.
- Do not mix or substitute reagents with those from other lots or sources.
- If samples generate values higher than the highest standard, dilute the samples with calibrator diluent and repeat the assay.
- Any variation in diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.
- Variations in sample collection, processing, and storage may cause sample value differences.
- This assay is designed to eliminate interference by other factors present in biological samples. Until all factors have been tested in the Quantikine[®] Immunoassay, the possibility of interference cannot be excluded.

TECHNICAL HINTS

- When mixing or reconstituting protein solutions, always avoid foaming.
- To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
- For best results, pipette reagents and samples into the center of each well.
- It is recommended that the samples be pipetted within 15 minutes.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- Substrate Solution should remain colorless until added to the plate. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
- Stop Solution should be added to the plate in the same order as the Substrate Solution. The color developed in the wells will turn from blue to yellow upon addition of the Stop Solution.

MATERIALS PROVIDED & STORAGE CONDITIONS

			STORAGE OF OPENED/
PART	PART #	DESCRIPTION	RECONSTITUTED MATERIAL
Mouse Fas Ligand	891009	96 well microplate (12 strips of 8 wells)	Return unused wells to the foil pouch containing
Microplate		coated with a monoclonal antibody	the desiccant pack. Reseal along entire edge of the
		specific for mouse Fas Ligand.	zip-seal. May be stored for up to 1 month at 2-8 °C.*
Mouse Fas Ligand	891011	2 vials of recombinant mouse Fas Ligand in	
Standard		a buffered protein base with preservatives;	
		lyophilized. <i>Refer to the vial label for</i>	
		reconstitution volume.	Use a new standard and control for each assay.
Mouse Fas Ligand	891012	2 vials of recombinant mouse Fas Ligand in	Discard within 8 hours of reconstitution.
Control		a buffered protein base with preservatives;	
		lyophilized. The assay value of the control	
		the label	
Mouco Eac Ligand	001010	12 ml of a natural antibody enorific	
Mouse ras Ligaliu	091010	for mouse East Ligand conjugated to	
Conjugate		horseradish peroxidase with preservatives	
Assav Diluent	895574	12 ml of a buffered protein base with	
RD1-47	0,5521	preservatives.	
Calibrator Diluent	895175	21 mL of a buffered protein base with	
RD5T		preservatives.	
Wash Buffer	895003	21 mL of a 25-fold concentrated solution	May be stored for up to 1 month at 2-8 °C.*
Concentrate		of buffered surfactant with preservative.	
		May turn yellow over time.	
Color Reagent A	895000	12 mL of stabilized hydrogen peroxide.	
Color Reagent B	895001	12 mL of stabilized chromogen	
		(tetramethylbenzidine).	
Stop Solution	895174	23 mL of diluted hydrochloric acid.	
Plate Sealers	N/A	4 adhesive strips.	

Store the unopened kit at 2-8 °C. Do not use past kit expiration date.

* Provided this is within the expiration date of the kit.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 500 mL graduated cylinder.
- Horizontal orbital microplate shaker (0.12" orbit) capable of maintaining a speed of 500 ± 50 rpm.
- Polypropylene test tubes for dilution of standards.

PRECAUTIONS

The Stop Solution provided with this kit is an acid solution.

Some components in this kit contain a preservative which may cause an allergic skin reaction. Avoid breathing mist.

Color Reagent B may cause skin, eye, and respiratory irritation. Avoid breathing fumes.

Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling. Refer to the SDS on our website prior to use.

SAMPLE COLLECTION & STORAGE

The sample collection and storage conditions listed below are intended as general guidelines. Sample stability has not been evaluated.

Cell Culture Supernates - Remove particulates by centrifugation. Assay immediately or aliquot and store samples at \leq -20 °C. Avoid repeated freeze-thaw cycles.

Serum - Allow blood samples to clot for 2 hours at room temperature before centrifuging for 20 minutes at 2000 x g. Remove serum and assay immediately or aliquot and store samples at \leq -20 °C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA as a anticoagulant. Centrifuge for 20 minutes at 2000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at \leq -20 °C. Avoid repeated freeze-thaw cycles.

Note: Heparin and citrate plasma have not been validated for use in this assay.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Mouse Fas Ligand Control - Reconstitute the control with 1.0 mL of deionized or distilled water. Mix thoroughly. Assay the control undiluted.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 20 mL of Wash Buffer Concentrate to 480 mL of deionized or distilled water to prepare 500 mL of Wash Buffer.

Substrate Solution - Color Reagents A and B should be mixed together in equal volumes within 15 minutes of use. Protect from light. 100 μL of the resultant mixture is required per well.

Mouse Fas Ligand Standard - Refer to the vial label for reconstitution volume.

Reconstitute the Mouse Fas Ligand Standard with Calibrator Diluent RD5T. Do not substitute other diluents. This reconstitution produces a stock solution of 2000 pg/mL. Allow the standard to sit for a minimum of 5 minutes with gentle mixing prior to making dilutions.

Use polypropylene tubes. Pipette 200 µL of Calibrator Diluent RD5T into each tube. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The undiluted Mouse Fas Ligand Standard (2000 pg/mL) serves as the high standard. Calibrator Diluent RD5T serves as the zero standard (0 pg/mL).



ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that all standards, control, and samples be assayed in duplicate.

- 1. Prepare all reagents, standard dilutions, control, and samples as directed in the previous sections.
- 2. Remove excess microplate strips from the plate frame, return them to the foil pouch containing the desiccant pack, and reseal.
- 3. Add 50 μL of Assay Diluent RD1-47 to each well.
- 4. Add 50 μL of standard, control, or sample per well. Cover with the adhesive strip provided. Incubate for 2 hours at room temperature on a horizontal orbital microplate shaker (0.12" orbit) set at 500 ± 50 rpm. A plate layout is provided to record standards and samples assayed.
- 5. Aspirate each well and wash, repeating the process four times for a total of five washes. Wash by filling each well with Wash Buffer (400 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
- 6. Add 100 μ L of Mouse Fas Ligand Conjugate to each well. Cover with a new adhesive strip. Incubate for 2 hours at room temperature on the shaker.
- 7. Repeat the aspiration/wash as in step 5.
- 8. Add 100 μ L of Substrate Solution to each well. Incubate for 30 minutes at room temperature **on the benchtop. Protect from light.**
- 9. Add 100 μ L of Stop Solution to each well. Gently tap the plate to ensure thorough mixing.
- 10. Determine the optical density of each well within 30 minutes, using a microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density (O.D.).

Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the mouse Fas Ligand concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

TYPICAL DATA

This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.



0.D.	Average	Corrected
0.063	0.064	
0.066		
0.100	0.100	0.036
0.101		
0.133	0.136	0.072
0.138		
0.203	0.206	0.142
0.208		
0.368	0.370	0.306
0.371		
0.637	0.646	0.582
0.656		
1.210	1.214	1.150
1.217		
1.938	2.022	1.958
2.107		
	O.D. 0.063 0.100 0.101 0.133 0.138 0.203 0.208 0.368 0.371 0.637 0.656 1.210 1.217 1.938 2.107	0.D. Average 0.063 0.064 0.066

PRECISION

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Three samples of known concentration were tested in twenty separate assays to assess inter-assay precision. Assays were performed by at least three technicians using two lots of components.

	Intra-Assay Precision			Inter-Assay Precision		
Sample	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean (pg/mL)	84	261	675	76	243	663
Standard deviation	8.1	9.0	20	6.8	12	33
CV (%)	9.6	3.4	3.0	8.9	4.9	5.0

RECOVERY

The recovery of mouse Fas Ligand spiked to three levels throughout the range of the assay in various matrices was evaluated.

Sample Type	Average % Recovery	Range
Cell culture supernates (n=6)	107	84-120%
Serum (n=6)	98	86-111%
EDTA plasma (n=6)	99	87-110%

LINEARITY

To assess the linearity of the assay, samples containing and/or spiked with high concentrations of mouse Fas Ligand were serially diluted with calibrator diluent to produce samples with values within the dynamic range of the assay.

		Cell culture supernates (n=6)	Serum (n=6)	EDTA plasma (n=6)
1.0	Average % of Expected	100	104	102
1:2	Range (%)	97-103	101-109	101-104
1:4	Average % of Expected	102	107	107
	Range (%)	96-107	101-115	104-110
1:8	Average % of Expected	104	106	108
	Range (%)	96-110	99-115	104-118
1:16	Average % of Expected	103	99	104
	Range (%)	93-108	90-114	93-115

SENSITIVITY

Two assays were evaluated and the minimum detectable dose (MDD) of mouse Fas Ligand ranged from 3.1-3.9 pg/mL. The mean MDD was 3.6 pg/mL.

The MDD was determined by adding two standard deviations to the mean O.D. value of twenty zero standard replicates and calculating the corresponding concentration.

CALIBRATION

This immunoassay is calibrated against highly purified NSO-expressed recombinant mouse Fas Ligand produced at R&D Systems[®].

SAMPLE VALUES

Serum/Plasma - Samples were evaluated for the presence of mouse Fas Ligand in this assay.

Sample Type	Mean of Detectable (pg/mL)	% Detectable	Range (pg/mL)
Serum (n=20)	83	15%	ND-138
EDTA plasma (n=20)	73	50%	ND-231

ND=Non-detectable

Cell Culture Supernates:

Mouse thymocytes (from one mouse thymus) were cultured for 3 days in RPMI supplemented with 10% fetal bovine serum and 50 μ M β -mercaptoethanol and stimulated with 10 ng/mL recombinant human IL-2. The plate in which the cells grew was pre-treated with 10 μ g/mL of hamster anti-mouse CD3. An aliquot of the cell culture supernate was removed, assayed for mouse Fas Ligand, and measured 243 pg/mL.

Mouse splenocytes (1 x 10⁶ cells/mL) were cultured for 3 days in RPMI supplemented with 10% fetal bovine serum and 50 μ M β -mercaptoethanol, and stimulated with 10 ng/mL recombinant human IL-2. An aliquot of the cell culture supernate was removed, assayed for mouse Fas Ligand, and measured 133 pg/mL.

SPECIFICITY

This assay recognizes natural and recombinant mouse Fas Ligand.

The factors listed below were prepared at 50 ng/mL in calibrator diluent and assayed for cross-reactivity. Preparations of the same factors at 50 ng/mL in a mid-range mouse Fas Ligand control were assayed for interference. No significant cross-reactivity or interference was observed.

Recombinant mouse:

C10	IL-17
Eotaxin	IL-18
E-Selectin	JE/MCP-1
Fas	KC
Flt-3 Ligand	L-Selectin
G-CSF	Leptin
GM-CSF	LIF
IFN-γ	MARC
IL-1a	MCP-5
IL-1β	MIP-1a
IL-2	MIP-2
IL-3	OPG
IL-4	OSM
IL-5	RANK Ligand
IL-6	RANTES
IL-7	TARC
IL-9	TNF-α
IL-10	TNF RI
IL-10 R	TNF RII
IL-12 p70	Тро
IL-13	VEGF R2

Recombinant human:

Fas Fas Ligand

REFERENCES

- 1. Walczak, H. and P.H. Krammer (2000) Exp. Cell Res. 256:58.
- 2. Mogi, M. et al. (2001) Lab. Invest. 81:177.
- 3. Dockrell, D.H. et al. (1998) J. Clin. Invest. 101:2394.
- 4. Suda, T. *et al*. (1995) J. Immunol. **154**:3806.
- 5. Suzuki, I. and P.J. Fink (2000) Proc. Natl. Acad. Sci. USA 97:1707.
- 6. Serrao, K.L. et al. (2001) Am. J. Physiol. 280:L298.
- 7. Liles, W.C. et al. (1996) J. Exp. Med. **184**:429.
- 8. Francavilla, S. et al. (2000) J. Clin. Endocrinol. Metab. 85:2692.
- 9. Giordano, C. *et al.* (1997) Science **275**:960.
- 10. Stassi, G. *et al.* (2000) Nature Immunol. **1**:483.
- 11. Saitoh, A. et al. (2000) J. Invest. Dermatol. 115:154.
- 12. Badie, B. *et al*. (2000) Exp. Neurol. **162**:290.
- 13. Griffith, T.S. *et al*. (1995) Science **270**:1189.
- 14. Lorz, C. *et al.* (2000) J. Am. Soc. Nephrol. **11**:1126.
- 15. Brown, S.B. and J. Savill (1999) J. Immunol. 162:480.
- 16. Song, J. et al. (2000) J. Clin. Invest. 106:1209.
- 17. Lu, L. *et al*. (1997) J. Immunol. **158**:5676.
- 18. O'Connell, P.J. et al. (2000) J. Immunol. 165:795.
- 19. Nagata, S. and P. Golstein (1995) Science 267:1449.
- 20. Lynch, D.H. *et al.* (1994) Immunity **1**:131.
- 21. Peitsch, M.C. and J. Tschopp (1995) Mol. Immunol. 32:761.
- 22. Hahne, M. *et al.* (1995) Int. Immunol. **7**:1381.
- 23. Takahashi, T. *et al*. (1994) Cell **76**:969.
- 24. Schneider, P. et al. (1998) J. Exp. Med. 187:1205.
- 25. Tanaka, M. *et al*. (1998) Nature Med. **4**:31.
- 26. Ayroldi, E. *et al.* (1999) Blood **94**:3456.
- 27. Tanaka, M. et al. (1995) EMBO J. 14:1129.
- 28. Takahashi, T. *et al.* (1994) Int. Immunol. **6**:1567.
- 29. Suda, T. *et al*. (1993) Cell **75**:1169.
- 30. Watanabe-Fukunaga, R. *et al*. (1992) J. Immunol. **148**:1274.
- 31. Pitti, R.M. *et al.* (1998) Nature **396**:699.
- 32. Raoul, C. et al. (2000) Curr. Opin. Neurobiol. 10:111.
- 33. Thilenius, A.R.B. et al. (1999) J. Immunol. 162:643.

PLATE LAYOUT

Use this plate layout to record standards and samples assayed.



NOTES

NOTES

All trademarks and registered trademarks are the property of their respective owners.

14

©2018 R&D Systems®, Inc.