

INTENDED USE & DESCRIPTION

For use as quantitative controls for the determination of cytokine concentrations in biological fluids. Concentrations have been assigned using R&D Systems® Quantikine kits. Controls are prepared in a buffered protein base with preservatives. They contain recombinant human cytokines at low, medium and high concentrations. Controls are supplied lyophilized.

STORAGE & STABILITY

Unreconstituted controls should be stored at 2-8 °C and are stable for at least 6 months from date of receipt. Depending on the analyte of interest, reconstituted controls may be stable when stored at ≤ -20 °C. Users should evaluate the frozen stability of the controls in their application or discard after use.

REAGENT PREPARATION

Reconstitute each vial with the volume of deionized or distilled water indicated in the chart below.

PROCEDURE & EXPECTED VALUES

Controls should be used undiluted and assayed as unknown specimens.

The acceptable ranges (±3 SD) for the analytes in these controls are printed below. Due to possible variations in techniques and methodologies, it is recommended that each laboratory determine its own target range. Laboratories using other test systems should establish their own acceptable ranges as these assays may produce different values.

Analyte	Catalog #	Kit Diluent	Water Recon. Volume	Units	Lot # P313599	Lot # P313602	Lot # P313605
Human Total Adiponectin/Acrp30	DRP300, SRP300, PDRP300	RD5-5	2.5 mL	ng/mL	11.8-26.7	34.3-63.8	71.6-118
		RD6-39			22.3-41.0	59.4-97.8	111-182
Human Leptin	DLP00, SLP00, PDLP00	RD5P (diluted 1:5)	2.5 mL	pg/mL	80.0-152	246-402	468-763
Human Resistin*	DRSN00, SRSN00, PDRSN00	RD5K	2.5 mL	ng/mL	0.462-1.07	1.31-2.93	3.85-7.19

TECHNICAL HINTS & LIMITATIONS OF THE PROCEDURE

- The ranges were determined using R&D Systems Quantikine kits. If expected values are not obtained, verify that the lot numbers on the vials correspond with the lot numbers listed above and the correct volume of deionized or distilled water was used for reconstitution of the controls.
- The results obtained with these controls depend upon several factors associated with methods and instrumentation. Test systems other than those supplied by R&D Systems may result in values that differ from those printed on this product datasheet.

**The purchase of this product conveys to the buyer the limited, non-exclusive, non-transferable right (without the right to resell, repackaging, or further sublicense) to use these reagents for non-commercial research purposes only. No other license is granted to the buyer whether expressly, by implication, by estoppel or otherwise. In particular, the purchase of this product does not include nor carry any right or license to use, develop, or otherwise exploit this product commercially, which includes without limitation, provision of services to a third party, generation of commercial databases, clinical diagnostics or therapeutics, or drug development. This product is manufactured under a license to U.S. Patent number 7,470,522. Any party desiring rights under this patent should contact Ryogen LLC, Montebello Park, 75 Montebello Road, Suffern, NY 10901.*