

INTENDED USE & DESCRIPTION

For use as quantitative controls for the determination of cytokine concentrations in biological fluids. Concentrations have been assigned using R&D Systems® Quantikine ELISA kits. Controls are prepared in a buffered protein base with preservatives. They contain human cytokines at low, medium and high concentrations. Controls are supplied lyophilized.

STORAGE & STABILITY

Unreconstituted controls should be stored at 2-8 °C and are stable for at least 6 months from date of receipt. Depending on the analyte of interest, reconstituted controls may be stable when stored at ≤ -20 °C. Users should evaluate the frozen stability of the controls in their application or discard after use.

PRECAUTION

The sTfR Controls contain human sTfR. This sTfR was tested at the donor level using an FDA licensed method and found to be non-reactive for anti-HIV-1/2 and Hepatitis B surface antigen. As no testing can offer complete assurance of freedom from infectious agents, these reagents should be handled as if capable of transmitting infection.

REAGENT PREPARATION

Reconstitute each vial with the volume of deionized or distilled water indicated in the chart below.

PROCEDURE & EXPECTED VALUES

Controls should be used undiluted and assayed as unknown specimens.

The acceptable ranges (±3 SD) for the analyte in these controls are printed below. Due to possible variations in techniques and methodologies, it is recommended that each laboratory determine its own target range. Laboratories using other test systems should establish their own acceptable ranges as these assays may produce different values.

Analyte	Catalog #	Kit Diluent	Water Recon. Volume	Lot # 1647241 (nmol/L)	Lot # 1647242 (nmol/L)	Lot # 1647243 (nmol/L)
Human sTfR	DTFRUO , STFRUO , PDTRUO	RD6-70	2.5 mL	3.80-6.20	14.7-24.1	22.4-36.6

TECHNICAL HINTS & LIMITATIONS OF THE PROCEDURE

- The ranges were determined using R&D Systems Quantikine ELISA kits. If expected values are not obtained, verify that the lot numbers on the vials correspond with the lot numbers listed above and the correct volume of deionized or distilled water was used for reconstitution of the controls.
- The results obtained with these controls depend upon several factors associated with methods and instrumentation. Test systems other than those supplied by R&D Systems may result in values that differ from those printed on this product datasheet.