



PRODUCT INFORMATION & MANUAL

Human VEGF Valukine™ ELISA

VAL106

For the quantitative determination of natural and recombinant
human VEGF concentrations

For research use only.
Not for diagnostic or therapeutic procedures.

Bio-Techne China Co. Ltd

P: +86 (21) 52380373 P: 8009881270 F: +86 (21) 52381001
info.cn@bio-techne.com

Please refer to the kit label for expiry date.
Novus kits are guaranteed for 3 months from date of receipt

Version202209.4

TABLE OF CONTENTS

I. BACKGROUND	2
II. OVERVIEW	3
III. ADVANTAGES	4
IV. EXPERIMENT	7
V. KIT COMPONENTS AND STORAGE	8
VI. PREPARATION	10
VII. ASSAY PROCEDURE	12
VIII. REFERENCES	14

I. BACKGROUND

Vascular endothelial growth factor (VEGF or VEGF-A), also known as vascular permeability factor (VPF), is a potent mediator of both angiogenesis and vasculogenesis in the fetus and adult (1-3). It is a member of the PDGF family that is characterized by the presence of eight conserved cysteine residues in a cystine knot structure and the formation of antiparallel disulfide-linked dimers (4). Humans express alternately spliced isoforms of 121, 145, 165, 183, 189, and 206 amino acids (aa) in length (4). VEGF165 appears to be the most abundant and potent isoform, followed by VEGF121 and VEGF189 (3, 4). Isoforms other than VEGF121 contain basic heparin-binding regions and are not freely diffusible (4). Human VEGF165 shares 88% aa sequence identity with corresponding regions of mouse and rat VEGF. VEGF is expressed in multiple cells and tissues including skeletal and cardiac muscle (5, 6), hepatocytes (7), osteoblasts (8), neutrophils (9), macrophages (10), keratinocytes (11), brown adipose tissue (12), CD34⁺ stem cells (13), endothelial cells (14), fibroblasts, and vascular smooth muscle cells (15). VEGF expression is induced by hypoxia and cytokines such as IL-1, IL-6, IL-8, oncostatin M and TNF- α (3, 4, 9, 16). VEGF isoforms are differentially expressed during development and in the adult (3).

VEGF dimers bind to two related receptor tyrosine kinases, VEGF R1 (also called Flt-1) and VEGF R2 (Flk-1/KDR), and induce their homodimerization and autophosphorylation (3, 4, 7, 17, 18). These receptors have seven extracellular immunoglobulin-like domains and an intracellular split tyrosine kinase domain. They are expressed on vascular endothelial cells and a range of non-endothelial cells. Although VEGF affinity is highest for binding to VEGF R1, VEGF R2 appears to be the primary mediator of VEGF angiogenic activity (3, 4). VEGF165 also binds the semaphorin receptor, neuropilin-1, which promotes complex formation with VEGF R2 (19).

VEGF is best known for its role in vasculogenesis. During embryogenesis, VEGF regulates the proliferation, migration, and survival of endothelial cells (3, 4), thus regulating blood vessel density and size, but playing no role in determining vascular patterns. VEGF promotes bone formation through osteoblast and chondroblast recruitment and is also a monocyte chemoattractant (20-22). After birth, VEGF maintains endothelial cell integrity and is a potent mitogen for micro- and macro-vascular endothelial cells. In adults, VEGF functions mainly in wound healing and the female reproductive cycle (3). In diseased tissues, VEGF promotes vascular permeability. It is thus thought to contribute to tumor metastasis by promoting both extravasation and tumor angiogenesis (23, 24). Various strategies have been employed therapeutically to antagonize VEGF-mediated tumor angiogenesis (25). Circulating VEGF levels correlate with disease activity in autoimmune diseases such as rheumatoid arthritis, multiple sclerosis and systemic lupus erythematosus (26).

II. OVERVIEW

A. PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for VEGF has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any VEGF present is bound by the immobilized antibody. After washing away any unbound substances, an enzyme-linked polyclonal antibody specific for VEGF is added to the wells. Following a wash to remove any unbound antibody-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of VEGF bound in the initial step. The color development is stopped and the intensity of the color is measured.

B. LIMITATIONS OF THE PROCEDURE

- ◆ **FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.**
- ◆ This kit is suitable for cell culture supernate, serum, EDTA Plasma and citrate plasma.
- ◆ The kit should not be used beyond the expiration date on the kit label.
- ◆ Do not mix or substitute reagents with those from other lots or sources.
- ◆ If samples generate values higher than the highest standard, dilute the samples with Diluent and repeat the assay.
- ◆ Any variation in operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

III. ADVANTAGES

A. PRECISION

Intra-assay Precision (Precision within an assay)

Three samples were tested twenty times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays)

Three samples were tested in twenty separate assays to assess inter-assay precision.

	Intra-assay Precision			Inter-assay Precision		
Sample	1	2	3	1	2	3
Mean (pg/mL)	128	401	810	129	388	749
Standard Deviation	8.24	26.0	40.2	12.0	31.1	70.2
CV%	6.4	6.5	5.0	9.3	8.0	9.4

B. RECOVERY

The recovery of human VEGF spiked to levels throughout the range of the assay in various matrices was evaluated.

Sample Type	Average % Recovery	Range
Cell culture media (n=4)	121	113 - 135%
Serum (n=3)	124	116 - 127%
Plasma (n=6)	100	78 - 119%

C. SENSITIVITY

The minimum detectable dose (MDD) of human VEGF is typically less than 7.8 pg/mL.

The MDD was determined by adding two standard deviations to the mean optical density value of twenty zero standard replicates and calculating the corresponding concentration.

D. CALIBRATION

This immunoassay is calibrated against highly purified *Sf-21*-expressed recombinant human VEGF produced at R&D Systems.

E. LINEARITY

To assess the linearity of the assay, samples were spiked with high concentrations of human VEGF in various matrices and diluted with Diluent 1× to produce samples with values within the dynamic range of the assay.

Dilution		Cell culture media (n=4)	Serum (n=3)	Plasma (n=6)
1:2	Average % of Expected	100	88	102
	Range (%)	96 - 104	87 - 89	87 - 110
1:4	Average % of Expected	101	91	108
	Range (%)	95 - 108	90 - 91	81 - 124
1:8	Average % of Expected	102	89	117
	Range (%)	94 - 111	88 - 91	106 - 125
1:16	Average % of Expected	101	80	102
	Range (%)	88 - 110	74 - 86	88 - 119

F. SAMPLE VALUES

Cell Culture Supernates - Human peripheral blood mononuclear cells (1×10^6 cells/mL) were cultured in RPMI supplemented with 5% fetal calf serum, 50 μ M β -mercaptoethanol, 2 mM L-glutamine, 100 U/mL penicillin, and 100 μ g/mL streptomycin sulfate. Cells were cultured unstimulated or stimulated 10 μ g/mL PHA. Aliquots of the cell culture supernates was removed and assayed for levels of natural VEGF and measured 332 pg/ml and 1440 pg/ml respectively.

Serum - Three serum samples were evaluated for the presence of human VEGF in this assay. All samples measured ranged from 217.5 to 375.5 pg/mL with an average of 282.4 pg/mL.

Plasma - Six human plasma samples were evaluated for the presence of human VEGF in this assay. All samples measured below the lowest standard, 15.6 pg/mL.

G. SPECIFICITY

This assay recognizes both natural and recombinant human VEGF. The following factors were prepared at 50 ng/mL and assayed for cross-reactivity. Preparations of the following factors at 50 ng/mL in a mid-range rhVEGF control were assayed for interference. No significant cross-reactivity or interference was observed.

Recombinant human				Recombinant mouse	
ANG	sgp130	IL-8	PDGF-AA	IL-1 α	IL-13
AR	GRO α	IL-10	PDGF-AB	IL-1 β	MIP-1 α
CNTF	GRO β	IL-11	PDGF-BB	IL-3	MIP-1 β
β -ECGF EGF	GRO γ	IL-12	P/GF	IL-4	P/GF-2
Epo	HB-EGF	IL-13	TGF- β 1	IL-5	TNF- α
FGF acidic	HGF	LAP (TGF- β 1)	TNF- α	IL-6	VEGF120
FGF basic	IL-1 α	M-CSF	TNF R	IL-9	VEGF164
FGF-4	IL-1 β	MCP-1	sTNF RII	IL-10	VEGF-R3
FGF-5	IL-2	MIP-1 α	VEGF165/P/GF		
FGF-6	IL-3	MIP-1 β	VEGF-B167		
FGF-7	IL-4	β -NGF	VEGF-C	Recombinant rat	
G-CSF	IL-5	OSM	VEGF-D	VEGF	
GM-CSF	IL-6	PD-ECGF	VEGF-R3		


VEGF-related factors showing cross-reactivity or interference:

- ◆ rhVEGF R1/Fc Chimera interfered at levels >500 pg/mL
- ◆ rhVEGF R2/Fc Chimera interfered at levels >2000 pg/mL
- ◆ rmVEGF R1 interfered at levels >500 pg/mL
- ◆ rmVEGF R2 interfered at levels >4000 pg/mL
- ◆ Canine VEGF showed 67% cross-reactivity

IV. EXPERIMENT

EXAMPLE STANDARD

The standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.



pg/mL	OD	Average	Corrected
0	0.013 0.016	0.015	-
15.63	0.036 0.035	0.036	0.021
31.25	0.059 0.062	0.061	0.046
62.5	0.131 0.115	0.123	0.109
125	0.283 0.222	0.253	0.238
250	0.471 0.531	0.501	0.487
500	0.995 1.024	1.010	0.995
1000	2.041 2.006	2.024	2.009

V. KIT COMPONENTS AND STORAGE

A. MATERIALS PROVIDED

Parts	Description	Size
Human VEGF Microplate	96 well polystyrene microplate (12 strips of 8 wells) coated with a mouse monoclonal antibody against human VEGF	1 plate
Human VEGF Conjugate	solution of polyclonal antibody against human VEGF conjugated to horseradish peroxidase	1 vial
Human VEGF Standard	recombinant human VEGF in a buffered protein base; lyophilized	2 vials
Calibrator Diluent (5×)	a 5× concentrated buffered protein base	1 vial
Wash Buffer Concentrate (25×)	a 25× concentrated solution of buffered surfactant	1 vial
TMB Substrate	TMB ELISA Substrate Solution	2 vials
Stop Solution	2 N sulfuric acid	1 vial
Plate Covers	adhesive strip	3 strips

B. STORAGE

Unopened Kit	Store at 2-8°C. Do not use past kit expiration date.	
Opened/ Reconstituted Reagents	Diluted Wash Solution	May be stored for up to 1 month at 2-8°C.*
	Stop Solution	
	Diluent 1×	
	Conjugate	
	TMB Substrate	
	Standard	Discard the VEGF stock solution and dilutions after 4 hours. Use a fresh standard for each assay.
Microplate Wells	Return unused wells to the foil pouch containing the desiccant pack, reseal along entire edge of zip-seal. May be stored for up to 1 month at 2-8°C.*	

* Provided this is within the expiration date of the kit.

C. OTHER SUPPLIES REQUIRED

- ◆ Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- ◆ Pipettes and pipette tips.
- ◆ Deionized or distilled water.
- ◆ Squirt bottle, manifold dispenser, or automated microplate washer.
- ◆ 500 mL graduated cylinder.

D. PRECAUTION

- ◆ The Stop Solution provided with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.
- ◆ VEGF is detectable in saliva. Take the necessary precautions to prevent contamination of the kit reagents.

VI. PREPARATION

A. SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤ -20 °C. Avoid repeated freeze-thaw cycles. Samples may require dilution with Diluent 1 \times .

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes at room temperature before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20 °C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20 °C. Avoid repeated freeze-thaw cycles.

B. SAMPLE PREPARATION

Serum samples require a 5-fold dilution. A suggested 5-fold dilution is 40 μ L of sample + 160 μ L of Diluent (1 \times).

Plasma samples require a 2-fold dilution. A suggested 2-fold dilution is 100 μ L of sample + 100 μ L of Diluent (1 \times).

C. REAGENT PREPARATION

Note: Bring all reagents to room temperature before use.

Wash Solution - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate (25 \times) into deionized or distilled water to prepare 500 mL of Wash Buffer.


Diluent 1 \times - Add 20 mL of Calibrator Diluent 5 \times into 80 mL of deionized or distilled water to prepare 100 mL of Diluent 1 \times .

VEGF Standard - Refer to the vial label for the reconstitution volume*. This reconstitution produces a stock solution of 2000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.

*if you have any question, please seek help from our Technical Support.

Pipette 500 μ L of Diluent 1 \times into each tube. Use the stock solution to produce a

dilution series (below). Mix each tube thoroughly before the next transfer. The 1000 pg/mL standard serves as the high standard. Diluent 1× serves as the zero standard (0 pg/mL).



D. TECHNICAL HINTS

- When mixing or reconstituting protein solutions, always avoid foaming.
- To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
- It is recommended that the samples be pipetted within 15 minutes.
- To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.
- TMB Substrate should remain colorless until added to the plate. Keep Substrate Solution protected from light. Substrate Solution should change from colorless to gradations of blue.
- Stop Solution should be added to the plate in the same order as the Substrate Solution. The color developed in the wells will turn from blue to yellow upon addition of the Stop Solution. Wells that are green in color indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution.

VII. ASSAY PROCEDURE

Note: Bring all reagents and samples to room temperature before use. It is recommended that all samples and standards be assayed in duplicate.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the foil pouch containing the desiccant pack, and reseal.
3. Add 100 μ L of Standard, sample, or control per well. Cover with the adhesive strip provided. Incubate for 2 hours at room temperature. A plate layout is provided for a record of standards and samples assayed.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Solution (400 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 200 μ L of human VEGF Conjugate to each well. Cover with a new adhesive strip. Incubate for 2 hours at room temperature.
6. Repeat the aspiration/wash as in step 4.
7. Add 200 μ L of TMB Substrate to each well. Incubate for 30 minutes at room temperature. **Protect from light.**
8. Add 50 μ L of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
9. Determine the optical density of each well within 10 minutes, using a microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.
10. **CALCULATION OF RESULTS:** Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Create

a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the VEGF concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

VIII. REFERENCES

1. Leung, D.W. et al. (1989) *Science* 246:1306.
2. Keck, P.J. et al. (1989) *Science* 246:1309.
3. Byrne, A.M. et al. (2005) *J. Cell. Mol. Med.* 9:777.
4. Robinson, C.J. and Stringer, S.E. (2001) *J. Cell. Sci.* 114:853.
5. Richardson, R.S. et al. (1999) *Am. J. Physiol.* 277:H2247.
6. Sugishita, Y. et al. (2000) *Biochem. Biophys. Res. Commun.* 268:657.
7. Yamane, A. et al. (1994) *Oncogene* 9:2683.
8. Goad, D.L. et al. (1996) *Endocrinology*137:2262.
9. Gaudry, M. et al. (1997) *Blood* 90:4153.
10. McLaren, J. et al. (1996) *J. Clin. Invest.* 98:482.
11. Diaz, B.V. et al. (2000) *J. Biol. Chem.* 275:642.
12. Asano, A. et al. (1997) *Biochem. J.* 328:179.
13. Bautz, F. et al. (2000) *Exp. Hematol.* 28:700.
14. Namiki, A. et al. (1995) *J. Biol. Chem.* 270:31189.
15. Nauck, M. et al. (1997) *Am. J. Respir. Cell. Mol. Biol.* 16:398.
16. Angelo, L.S. and R. Kurzrock (2007) *Clin. Cancer Res.* 13:2825.
17. Neufeld, G. et al. (1999) *FASEB. J.* 13:9.
18. Kowalewski, M.P. et al. (2005) Accession #ABB82619.
19. Pan, Q. et al. (2007) *J. Biol. Chem.* 282:24049.
20. Weis, S.M. and D.A. Cheresh (2005) *Nature* 437:497.
21. Breier, G. (2000) *Semin. Thromb. Hemost.* 26:553.
22. Barleon, B. et al. (1996) *Blood* 87:3336.
23. Weis, S.M. and D.A. Cheresh (2005) *Nature* 437:497.
24. Thurston, G. (2002) *J. Anat.* 200:575.
25. Grothey, A. and E. Galanis (2009) *Nat. Rev. Clin. Oncol.* 6:507.
26. Carvalho, J.F. et al. (2007) *J. Clin. Immunol.* 27:246.

PLATE LAYOUT

Use this plate layout to record standards and samples assayed.

1								
2								
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
	A	B	C	D	E	F	G	H



产品信息及操作手册

人 VEGF Valukine™ ELISA 试剂盒

目录号: VAL106

适用于定量检测天然和重组人血管内皮生长因子 (VEGF) 的含量

科研专用, 不可用于临床诊断

Bio-Techne China Co. Ltd

P: +86 (21) 52380373 P: 8009881270 F: +86 (21) 52381001

info.cn@bio-techne.com

有效期详见试剂盒包装标签

Novus 试剂盒确保在你收货日期 3 个月内有效

目录

I. 背景.....	18
II. 概述.....	19
III. 优势.....	20
IV. 实验标准.....	23
V. 试剂盒组成及储存.....	24
VI. 实验前准备.....	26
VII. 操作步骤.....	28
VIII. 参考文献.....	29

I. 背景

血管内皮生长因子（VEGF或VEGF-A）又叫血管通透因子，是一种针对胎儿和成人的血管新生和血管生成的有效调节因子（1-3）。血管内皮生长因子属于PDGF家族，该家族蛋白的特点是在胱氨酸结和通过反向平行二硫键结合的二聚体的结构处，存在8个保守的胱氨酸残基。经过选择性剪切和氨基酸的序列长度，人类表达不同的亚型，包括：VEGF121、VEGF145、VEGF165、VEGF183、VEGF189和VEGF206等等（4）。

其中VEGF165是表达量最高的一种亚型，其次是VEGF121和VEGF189（3, 4）。除了VEGF121，其它的亚型都包含了基本的肝素结合区域且不能自由扩散（4）。人类的VEGF165与小鼠和大鼠的相应的蛋白氨基酸序列同源性为88%。VEGF在各种细胞和组织中表达，包括骨骼肌细胞和心肌细胞（5, 6）、肝细胞（7）、成骨细胞（8）、中性粒细胞（9）、巨噬细胞（10）、角质形成细胞（11）、棕色脂肪细胞（12）、CD34⁺干细胞（13）、内皮细胞（14）、成纤维细胞、血管平滑肌细胞等等（15）。VEGF的表达收到缺氧和细胞因子的诱导，包括IL-1、IL-6、IL-8、抑瘤素M和肿瘤坏死因子等（3, 4, 9, 16）。在发育过程和成体中，VEGF的表达量也是不同的（3）。

VEGF的二聚体与2个相关的酪氨酸激酶受体相结合，即VEGF R1（也叫Flt-1）和VEGF R2（Flk-1/KDR）。VEGF同时可以诱导后者的同型二聚体和自磷酸化过程（3, 4, 7, 17, 18）。这些受体拥有7个胞外的免疫球蛋白的域和一个胞内分开的酪氨酸受体域。血管内皮细胞和其它一些非内皮细胞都表达VEGF受体。尽管VEGF与VEGF R1的亲合力最高，但VEGF R2 却是调控VEGF的血管生成的主要因子（3, 4）。VEGF165也结合臂板蛋白（semaphorin）受体，神经纤毛蛋白1（neuropilin-1），由此促进与VEGFR2的复合体形成（19）。

VEGF因其参与血管生成而著称。在胚胎发育过程中，VEGF调控内皮细胞的增殖、迁移和生存（3, 4），并由此调控血管的密度、体积；但对于血管形成的格局并不起作用。VEGF通过成骨细胞和软骨细胞的招募促进骨骼的形成，它同时也是一个单核细胞趋化因子（20-22）。在产后，VEGF维持血管内皮细胞的完整性，并且是大/小血管内皮细胞的有效丝裂原。在成体中，VEGF主要在伤口修复和女性的生殖周期发挥作用（3）。在疾病组织当中，VEGF促进血管的通透性。因此，VEGF通过外渗和肿瘤血管生成参与了肿瘤的转移过程（23, 24）。针对阻断VEGF活性的各种治疗策略正在被用于控制由VEGF诱导的肿瘤血管生成（25）。体内循环的VEGF水平与自身免疫性疾病（如类风湿关节炎、多发性硬化症、系统性红斑狼疮等）的病症程度相关（26）。

II. 概述

A. 检测原理

本实验采用双抗体夹心ELISA法。抗人VEGF单抗包被于微孔板上，样品和标准品中的VEGF会与固定在板上的抗体结合，游离的成分被洗去；加入辣根过氧化酶标记的抗人VEGF多抗，与结合在微孔板上的VEGF结合而形成免疫复合物，游离的成分被洗去；加入TMB底物溶液，溶液颜色逐渐变成蓝色，加入终止液溶液变黄并且停止变化。用酶标仪测定吸光度。

B. 检测局限

- ◆ 仅供科研使用，不可用于体外诊断；
- ◆ 该试剂盒适用于细胞培养上清样本，血清样本，EDTA血浆样本和枸橼酸钠血浆样本；
- ◆ 请在试剂盒有效期内使用；
- ◆ 不同试剂盒及不同批号试剂盒的组分不能混用；
- ◆ 样本值若大于标准曲线的最高值，应将样本用稀释剂（1×）稀释后重新检测；
- ◆ 检测结果的不同可由多种因素引起，包括实验人员的操作、移液器的使用方式、洗板技术、反应时间或温度、试剂盒的储存等。

III. 优势

A. 精确度

板内精确度（同一板内不同孔间的精确度）

已知浓度的三个样本，在同一板内分别检测20次，以确定板内精确度。

板间精确度（不同板之间的精确度）

已知浓度的三个样本，在不同板间分别检测20次，以确定板间精确度。

样本	板内精确度			板间精确度		
	1	2	3	1	2	3
平均值 (pg/mL)	128	401	810	129	388	749
标准差	8.24	26.0	40.2	12.0	31.1	70.2
CV%	6.4	6.5	5.0	9.3	8.0	9.4

B. 回收率

在不同类型样本中掺入检测范围内不同水平的人VEGF，测定其回收率。

样本类型	平均回收率	范围
细胞培养上清 (n=4)	121	113 - 135%
血清 (n=3)	124	116 - 127%
血浆 (n=6)	100	78 - 119%

C. 灵敏度

人VEGF的最低可测剂量（MDD）一般小于7.8pg/mL。

MDD是根据20个重复的零标准品孔的吸光度值的平均值加两倍标准差计算得到的相对应浓度。

D. 校正

此ELISA试剂盒经由R&D Systems生产的Sf-21昆虫细胞系表达的高纯度重组人VEGF蛋白所校正。

E. 线性

在不同类型样本中掺入高浓度的人VEGF，然后用稀释剂将样本稀释到检测范围内，测定其线性。

稀释倍数		细胞培养上清 (n=4)	血清 (n=3)	血浆 (n=6)
1:2	平均值/期待值 (%)	100	88	102
	范围 (%)	96 - 104	87 - 89	87 - 110
1:4	平均值/期待值 (%)	101	91	108
	范围 (%)	95 - 108	90 - 91	81 - 124
1:8	平均值/期待值 (%)	102	89	117
	范围 (%)	94 - 111	88 - 91	106 - 125
1:16	平均值/期待值 (%)	101	80	102
	范围 (%)	88 - 110	74 - 86	88 - 119

F. 样本预值

细胞上清样本 - 人的外周血单核细胞 (1×10^6 细胞/mL) 培养于含有5%胎牛血清的RPMI1640培养基中，细胞培养基还含有2mM L-谷氨酰胺、50 μ M β -巯基乙醇、100U/mL青霉素, 100 μ g/mL 链霉素；不刺激细胞，或加10 μ g/mL PHA刺激细胞，取细胞上清液测定VEGF含量，分别测得332pg/ml和1440pg/ml。

血清样本 - 使用本试剂盒检测了3份人血清样本中VEGF的水平。3份样本的检测值在217.5 - 375.5pg/mL之间，平均值为282.4pg/mL。

血浆样本 - 使用本试剂盒检测了6份人血浆样本中VEGF的水平。6份样本的检测值均低于最低标准品，15.6pg/mL。

G. 特异性

此ELISA法可检测天然及重组人VEGF蛋白。将以下因子用稀释剂 ($1 \times$) 配置成50ng/mL的浓度来检测与人VEGF的交叉反应。将50ng/mL的干扰因子掺入中间范围的重组人VEGF对照品中，来检测对人VEGF的干扰。没有观察到明显的交叉反应或干扰。

重组人蛋白				重组小鼠蛋白	
ANG	sgp130	IL-8	PDGF-AA	IL-1 α	IL-13
AR	GRO α	IL-10	PDGF-AB	IL-1 β	MIP-1 α
CNTF	GRO β	IL-11	PDGF-BB	IL-3	MIP-1 β
β -ECGF EGF	GRO γ	IL-12	P/GF	IL-4	P/GF-2
Epo	HB-EGF	IL-13	TGF- β 1	IL-5	TNF- α
FGF acidic	HGF	LAP (TGF- β 1)	TNF- α	IL-6	VEGF120
FGF basic	IL-1 α	M-CSF	TNF R	IL-9	VEGF164
FGF-4	IL-1 β	MCP-1	sTNF RII	IL-10	VEGF-R3
FGF-5	IL-2	MIP-1 α	VEGF165/P/GF		
FGF-6	IL-3	MIP-1 β	VEGF-B167		
FGF-7	IL-4	β -NGF	VEGF-C	重组大鼠蛋白	
G-CSF	IL-5	OSM	VEGF-D	VEGF	
GM-CSF	IL-6	PD-ECGF	VEGF-R3		


与人VEGF有交叉反应或表现出干扰的因子：

- ◆ 重组人VEGF R1/Fc Chimera大于500pg/mL时有干扰
- ◆ 重组人VEGF R2/Fc Chimera大于2000pg/mL时有干扰
- ◆ 重组小鼠VEGF R1大于500pg/mL时有干扰
- ◆ 重组小鼠VEGF R2大于4000pg/mL时有干扰
- ◆ 与犬科VEGF有67%的交叉反应

IV. 实验标准

标准曲线实例

该标准曲线数据仅供参考，每次实验应绘制其对应的标准曲线。



V. 试剂盒组成及储存

A. 试剂盒组成

组成	描述	规格
人VEGF Microplate	包被抗体的96孔聚苯乙烯板，8孔×12条	1块板
人VEGF Conjugate	酶标检测人VEGF抗体	1瓶
人VEGF Standard	标准品（冻干）	2瓶
Calibrator Diluent（5×）	浓缩稀释剂（5×）	1瓶
Wash Buffer Concentrate（25×）	浓缩洗涤缓冲液（25×）	1瓶
TMB Substrate	TMB底物溶液	2瓶
Stop Solution	终止液	1瓶
Plate Covers	封板胶纸	3张

B. 试剂盒储存

未开封试剂盒	2-8℃储存；请在试剂盒有效期内使用	
已打开，稀释或重溶的试剂	洗涤缓冲液（1×）	2-8℃储存，最多30天*
	终止液	
	稀释剂（1×）	
	酶标检测抗体	
	TMB底物溶液	
	标准品	已重溶和稀释的标准品仅限4小时内使用。每次实验需新鲜配置标准品。
包被的微孔板条	将未用的板条放回带有干燥剂的铝箔袋内，密封；2-8℃储存，最多 30 天*	

*必须在试剂盒有效期内。

C. 实验所需自备试验器材

- ◆ 酶标仪（可测量450nm检测波长的吸收值及540nm或570nm校正波长的吸收值）
- ◆ 高精度加液器及一次性吸头
- ◆ 蒸馏水或去离子水
- ◆ 洗瓶（喷瓶）、多通道洗板器或自动洗板机
- ◆ 500mL量筒

D. 注意事项

- ◆ 试剂盒中的终止液是酸性溶液，使用时请做好眼睛、手、面部及衣服的保护。
- ◆ 人唾液中含有达到检测浓度的VEGF。实验时请戴口罩，防止唾液污染试剂。

VI. 实验前准备

A. 样品收集及储存

细胞培养上清液：颗粒物应离心去除；立刻检测样本。样本收集后若不及时检测，需按一次使用量分装，冻存于-20℃冰箱内，避免反复冻融。样本可能需要用稀释剂（1×）稀释。

血清样本：用血清分离管(SST)分离血清。使血样室温凝集30分钟，然后1000 x g离心15分钟。吸取血清样本之后即刻用于检测，或者分装，-20℃贮存备用。避免反复冻融。

血浆样本：使用EDTA或枸橼酸钠作为抗凝剂收集血浆。然后1000 x g离心15分钟，需在30分钟内收集血浆样本之后即刻用于检测，或者分装，-20℃贮存备用。避免反复冻融。

B. 样本准备工作

血清样本需要用稀释剂（1×）5倍稀释后进行检测，即40μL血清+160μL稀释剂（1×）。

血浆样本需要用稀释剂（1×）2倍稀释后进行检测，即100μL血浆+100μL稀释剂（1×）。

C. 检测前准备工作

使用前请将所有试剂放置于室温


洗涤液：从冰箱中取出的浓缩洗涤液可能有结晶，属于正常现象；放置室温，轻摇混匀，待结晶完成溶解后再配制洗涤液。可将20mL浓缩洗涤液用蒸馏水或去离子水稀释配置成500mL工作浓度的洗涤液。未用完的放回4℃。

稀释剂（1×）：可将20mL浓缩稀释剂用80mL蒸馏水或去离子水稀释配置成100mL工作浓度的稀释剂。

标准品：重溶冻干标准品，**重溶体积请参考瓶身标签**，得到浓度为2000pg/mL标准品母液。轻轻震荡至少15分钟，其充分溶解。

*如有疑问，请咨询我们的技术支持。

每个稀释管中加入500μL稀释剂（1×）。将标准品母液参照下图做系列稀释，每管须充分混匀后再移液到下一管。1000pg/mL的标准品可用作标准曲线最高点，稀释剂（1×）可用作标准曲线零点（0pg/mL）。



D. 技术小提示

- ◆ 当混合或重溶蛋白液时，尽量避免起沫；
- ◆ 为了避免交叉污染，配置不同浓度标准品、上样、加不同试剂都需要更换枪头。另外不同试剂请分别使用不同的移液槽；
- ◆ 建议15分钟内完成一块板的上样；
- ◆ 每次孵育时，正确使用封板胶纸可保证结果的准确性；
- ◆ TMB底物溶液在上板前应为无色，请避光保存；加入微孔板后，将由无色变成不同深度的蓝色；
- ◆ 终止液上板顺序应同TMB底物溶液上板顺序一致；加入终止液后，孔内颜色由蓝变黄；若孔内有绿色，则表明孔内液体未混匀请充分混合。

VII. 操作步骤

使用前请将所有试剂和样本放置于室温，建议所有的实验样本和标准品做复孔检测。

1. 按照上一节的说明，准备好所有需要的试剂和标准品；
2. 从已平衡至室温的密封袋中取出微孔板，未用的板条请放回铝箔袋内，重新封口；
3. 分别将不同浓度标准品，实验样本或者质控品加入相应孔中，每孔100 μ L。用封板胶纸封住反应孔，室温孵育2小时。说明书提供了一张96孔模板图，可用于记录标准品和试验样本的板内位置；
4. 将板内液体吸去，使用洗瓶、多通道洗板器或自动洗板机洗板。每孔加洗涤液400 μ L，然后将板内洗涤液吸去。重复操作4次。每次洗板尽量吸去残留液体会有助于得到好的实验结果。最后一次洗板结束，请将板内所有液体吸干或将板倒置，在吸水纸拍干所有残留液体；
5. 在每个微孔内加入200 μ L酶标检测抗体。用封板胶纸封住反应孔，室温孵育2小时；
6. 重复第4步洗板操作；
7. 在每个微孔内加入200 μ L TMB底物溶液，室温孵育30分钟。**注意避光；**
8. 在每个微孔内加入50 μ L终止液，孔内溶液颜色会从蓝色变为黄色。如果溶液颜色变为绿色或者颜色变化不一致，请轻拍微孔板，使溶液混合均匀；
9. 加入终止液后10分钟内，使用酶标仪测量450nm的吸光度值，设定540nm或570nm作为校正波长。如果没有使用双波长校正，结果准确度可能会受影响；
10. **计算结果：**将每个标准品和样品的校正吸光度值(OD₄₅₀-OD₅₄₀/OD₅₇₀)、复孔读数取平均值，然后减去平均零标准品OD值。使用计算机软件作四参数逻辑(4-PL)曲线拟合创建标准曲线。另一种方法是，可以通过绘制标准品浓度做对数与相应OD值对数生成曲线，并通过回归分析确定最佳拟合线。这个过程可生成一个足够使用但不太精确的数据拟合。若样本经过稀释，计算浓度时应乘以稀释倍数。

VIII. 参考文献

1. Leung, D.W. et al. (1989) *Science* 246:1306.
2. Keck, P.J. et al. (1989) *Science* 246:1309.
3. Byrne, A.M. et al. (2005) *J. Cell. Mol. Med.* 9:777.
4. Robinson, C.J. and Stringer, S.E. (2001) *J. Cell. Sci.* 114:853.
5. Richardson, R.S. et al. (1999) *Am. J. Physiol.* 277:H2247.
6. Sugishita, Y. et al. (2000) *Biochem. Biophys. Res. Commun.* 268:657.
7. Yamane, A. et al. (1994) *Oncogene* 9:2683.
8. Goad, D.L. et al. (1996) *Endocrinology*137:2262.
9. Gaudry, M. et al. (1997) *Blood* 90:4153.
10. McLaren, J. et al. (1996) *J. Clin. Invest.* 98:482.
11. Diaz, B.V. et al. (2000) *J. Biol. Chem.* 275:642.
12. Asano, A. et al. (1997) *Biochem. J.* 328:179.
13. Bautz, F. et al. (2000) *Exp. Hematol.* 28:700.
14. Namiki, A. et al. (1995) *J. Biol. Chem.* 270:31189.
15. Nauck, M. et al. (1997) *Am. J. Respir. Cell. Mol. Biol.* 16:398.
16. Angelo, L.S. and R. Kurzrock (2007) *Clin. Cancer Res.* 13:2825.
17. Neufeld, G. et al. (1999) *FASEB. J.* 13:9.
18. Kowalewski, M.P. et al. (2005) Accession #ABB82619.
19. Pan, Q. et al. (2007) *J. Biol. Chem.* 282:24049.
20. Weis, S.M. and D.A. Cheresh (2005) *Nature* 437:497.
21. Breier, G. (2000) *Semin. Thromb. Hemost.* 26:553.
22. Barleon, B. et al. (1996) *Blood* 87:3336.
23. Weis, S.M. and D.A. Cheresh (2005) *Nature* 437:497.
24. Thurston, G. (2002) *J. Anat.* 200:575.
25. Grothey, A. and E. Galanis (2009) *Nat. Rev. Clin. Oncol.* 6:507.
26. Carvalho, J.F. et al. (2007) *J. Clin. Immunol.* 27:246.

96 孔模板图

请使用 96 孔模板图来记录标准品及样本在板内的位置

